3.5g Sky Blue for Henry Schein Inc.

Henry Schein Inc. expects this glove to match PurAloe glove performance (softness and fitting), and the glove is expected to be softer Hence the modulus @ M300% should be ≤ 6 MPa. Henry Schein Inc. is fine to run Accelerator Free formulation but does not accept additional charges. Customer request Expected glove length for XS - S : Min 220mm and M - XL : Min 230mm. The glove should be tested for use with Chemotherapy Drugs and Fentanyl Citrate specifically. The glove should be free from holes as pin holes are not allowed in the glove. The colour of the glove is defined as sky blue.

Following are the additional details

Target Weight (g) / Thickness :

- XS : 2.8g +/- 0.2
- S : 3.0g +/- 0.2
- M : 3.5g +/- 0.2
- L : 4.0g +/- 0.2
- XL : 4.5g +/- 0.2

Cuff : Min 0.04
Palm : Min 0.06
Finger : Min 0.07

Length (mm) :

- XS - S : Min 220
- M - XL : Min 230

Glove width (mm) :

- XS : 70 +/- 10
- S : 80 +/- 10
- M : 95 +/- 10
- L : 110 +/- 10
- XL : 120 +/- 10

Tested for use with Chemotherapy Drugs and Fentanyl Citrate
Modulus @ M300% Elongation: ≤ 6 MPa
Freedom from holes: Pin hole is not allowed

Commercial Justification

We have maintained a strong working relationship with the customer in the CRSM and synthetic surgical product areas, with frequent communication between the media management teams of both parties. Recently, due to issues with CR material sourcing, we have taken the opportunity to explore and clarify opportunities in a new business area, the Nitrile XM Glove. We are targeting a monthly volume of 40 million pieces by the end of 2025 for standard Nitrile gloves. In fact, we

expect the total Nitrile volume to reach 30-40 million pieces per month starting in January. Additionally, we have completed the artwork for Puraloe and are currently developing monthly production for the 3.5g Sky Blue and 4.8g Blue gloves.